

Thursday, June 10, 2021, Time 14:17

Wintherm32v3 Version 3.31.143 Uni

Instrument: F200s

Instrument Program Version 290

Instrument Serial Number: 2520

Sample Name: A1-F4-20210610

Thickness: 37.89mm

Rear Left : 37.85 mm

Rear Right : 37.90 mm

Front Left: 37.90 mm

Front Right: 37.92 mm

[ ]

[ ]

Thickness obtained : from instrument

TEST RUN

Calibration used : User Type

Calibration File Id: IRMMCalibration\_sn2520

Number of transducers per plate: 1

Number of transducers used per plate: 1

Number of Setpoints: 3

Setpoint duration: 89 min

Block Averages for setpoint 1 in SI units

	Tupper [°C]	Tlower [°C]	Qupper [µV]	Qlower [µV]	Lambda [W/mK]	SampleU [°C]	SampleL [°C]
-ne-	16.05	24.60	-11863	4012	1.732	19.39	23.87
-ne-	15.00	24.96	-6737	2001	0.6134	17.39	24.38
-te-	15.00	25.01	-5425	2374	0.5215	16.96	24.30
-te-	15.00	25.01	-4850	2791	0.5062	16.75	24.17
-te-	15.00	25.01	-4500	3112	0.5028	16.63	24.06
-te-	15.00	25.01	-4274	3333	0.5015	16.54	24.00
-te-	15.00	25.01	-4126	3474	0.5003	16.49	23.95
-te-	15.00	25.01	-4019	3580	0.4998	16.45	23.92
-te-	15.00	25.01	-3961	3635	0.4991	16.42	23.90
-te-	15.01	25.01	-3909	3684	0.4987	16.41	23.89
-te-	15.01	25.01	-3883	3707	0.4983	16.39	23.88
-pe-	15.00	25.01	-3869	3720	0.4983	16.39	23.87
-pe-	15.00	25.01	-3852	3734	0.4979	16.38	23.87
-pe-	15.00	25.01	-3842	3743	0.4978	16.38	23.87
-pe-	15.00	25.01	-3834	3747	0.4974	16.38	23.87
-pe-	15.00	25.01	-3828	3751	0.4973	16.38	23.86
-pe-	15.00	25.01	-3826	3750	0.4972	16.38	23.87
-pe-	15.00	25.01	-3822	3754	0.4970	16.38	23.87
-pe-	15.00	25.01	-3820	3755	0.4969	16.37	23.86
-pe-	15.00	25.01	-3817	3756	0.4965	16.37	23.87
-pe-	15.01	25.01	-3819	3754	0.4966	16.37	23.87

Thursday, June 10, 2021, Time 15:46

Setpoint No. 1

Setpoint Upper: 15.00 °C

Setpoint Lower: 25.00 °C

Temperature Upper: 16.37 °C

CalibFactor Upper: 0.025935

Results Upper: 0.5008 W/mK

Temperature Lower: 23.87 °C

CalibFactor Lower: 0.025942

Results Lower: 0.4925 W/mK

Percent Difference: 1.66%

Thermal Equilibrium Criteria:

Temperature Equilibrium: 0.20

Between Block HFM Equil.: 49

HFM Percent Change: 2.00

Min Number of Blocks: 4

Calculation Blocks: 3

Setpoint duration: 97 min

Block Averages for setpoint 2 in SI units

	Tupper [°C]	Tlower [°C]	Qupper [μV]	Qlower [μV]	Lambda [W/mK]	SampleU [°C]	SampleL [°C]
-ne-	10.20	20.08	-11769	-5029	1.160	13.78	20.93
-te-	10.00	20.00	-7078	414	0.5103	12.66	19.93
-te-	10.00	20.00	-6006	1410	0.5003	12.25	19.59
-te-	10.00	20.00	-5257	2151	0.4966	11.97	19.36
-te-	10.00	20.00	-4741	2656	0.4941	11.79	19.20
-te-	10.00	20.00	-4401	3000	0.4929	11.66	19.09
-te-	10.00	20.01	-4178	3224	0.4920	11.57	19.02
-te-	10.00	20.00	-4023	3377	0.4912	11.52	18.97
-te-	10.00	20.01	-3921	3479	0.4908	11.48	18.94
-te-	10.00	20.00	-3856	3543	0.4905	11.45	18.91
-te-	10.01	20.01	-3814	3586	0.4903	11.44	18.90
-pe-	10.00	20.00	-3791	3608	0.4902	11.43	18.89
-pe-	10.00	20.00	-3764	3635	0.4901	11.41	18.88
-pe-	10.00	20.00	-3748	3650	0.4903	11.41	18.87
-pe-	10.00	20.00	-3735	3665	0.4900	11.41	18.88
-pe-	10.00	20.00	-3729	3670	0.4899	11.40	18.87
-pe-	10.00	20.00	-3728	3670	0.4897	11.40	18.88
-pe-	10.00	20.00	-3724	3673	0.4897	11.40	18.87
-pe-	10.00	20.00	-3721	3677	0.4899	11.40	18.87
-pe-	10.00	20.00	-3718	3679	0.4897	11.40	18.87
-pe-	10.00	20.01	-3719	3678	0.4896	11.40	18.87
-pe-	10.00	20.00	-3717	3679	0.4897	11.40	18.87
-pe-	10.01	20.01	-3718	3676	0.4895	11.40	18.87

Thursday, June 10, 2021, Time 17:24

Setpoint No. 2  
Setpoint Upper: 10.00 °C  
Setpoint Lower: 20.00 °C  
Temperature Upper: 11.40 °C  
CalibFactor Upper: 0.026229  
Results Upper: 0.4946 W/mK  
Temperature Lower: 18.87 °C  
CalibFactor Lower: 0.025977  
Results Lower: 0.4846 W/mK  
Percent Difference: 2.05%

Thermal Equilibrium Criteria:  
Temperature Equilibrium: 0.20  
Between Block HFM Equil.: 49  
HFM Percent Change: 2.00  
Min Number of Blocks: 4  
Calculation Blocks: 3

Setpoint duration: 97 min

Block Averages for setpoint 3 in SI units

	Tupper [°C]	Tlower [°C]	Qupper [μV]	Qlower [μV]	Lambda [W/mK]	SampleU [°C]	SampleL [°C]
-ne-	5.30	15.13	-11451	-4888	1.152	8.92	15.97
-ne-	4.95	14.92	-6958	321	0.5067	7.70	14.89
-te-	5.00	15.00	-5870	1397	0.4982	7.31	14.60
-te-	5.00	15.00	-5128	2113	0.4925	7.03	14.36
-te-	5.00	15.00	-4629	2604	0.4893	6.84	14.21
-te-	5.00	15.00	-4296	2938	0.4878	6.71	14.10
-te-	5.00	15.00	-4079	3162	0.4871	6.62	14.03
-te-	5.00	15.01	-3928	3314	0.4865	6.57	13.98
-te-	5.00	15.01	-3829	3409	0.4857	6.53	13.95
-te-	5.00	15.01	-3765	3478	0.4857	6.50	13.93
-te-	5.00	15.01	-3722	3522	0.4856	6.49	13.91
-pe-	5.00	15.00	-3701	3540	0.4855	6.48	13.90
-pe-	4.99	15.00	-3677	3565	0.4854	6.47	13.89
-pe-	4.99	15.00	-3661	3581	0.4852	6.46	13.89
-pe-	5.00	15.00	-3655	3589	0.4854	6.46	13.89
-pe-	5.00	15.00	-3648	3597	0.4853	6.46	13.88
-pe-	5.00	15.00	-3642	3603	0.4852	6.45	13.88
-pe-	5.00	15.00	-3644	3603	0.4852	6.45	13.89

-pe-	5.00	15.00	-3642	3605	0.4851	6.45	13.89
-pe-	5.00	15.00	-3640	3607	0.4851	6.45	13.89
-pe-	5.00	15.01	-3638	3608	0.4849	6.45	13.89
-pe-	5.00	15.00	-3638	3608	0.4850	6.45	13.89
-pe-	5.00	15.01	-3640	3605	0.4847	6.45	13.89

Thursday, June 10, 2021, Time 19:02

Setpoint No. 3  
Setpoint Upper: 5.00 °C  
Setpoint Lower: 15.00 °C  
Temperature Upper: 6.45 °C  
CalibFactor Upper: 0.026523  
Results Upper: 0.4917 W/mK  
Temperature Lower: 13.89 °C  
CalibFactor Lower: 0.026010  
Results Lower: 0.4780 W/mK  
Percent Difference: 2.83%

Thermal Equilibrium Criteria:  
Temperature Equilibrium: 0.20  
Between Block HFM Equil.: 49  
HFM Percent Change: 2.00  
Min Number of Blocks: 4  
Calculation Blocks: 3

Results Table -- SI Units

Mean Temp	Upper Cond	Lower Cond	Average Cond
20.12	0.5008	0.4925	0.4967
15.14	0.4946	0.4846	0.4896
10.17	0.4917	0.4780	0.4849